



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Megaquant D-Fructose/D-Glucose Assay Kit
Cat. No: K-FRGLMQ
Lot. No: 110101
Expiry Date: December, 2013

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 05/01/2011

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Megaquant D-Fructose/D-Glucose Assay Kit
Cat. No: K-FRGLMQ
Lot. No: 91019
Expiry Date: December, 2013

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 23/10/2009

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Megaquant D-Fructose/D-Glucose Assay Kit
Cat. No: K-FRGLMQ
Lot. No: 90803
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 21/08/2009

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Megaquant D-Fructose/D-Glucose Assay Kit
Cat. No: K-FRGLMQ
Lot. No: 90509
Expiry Date: December, 2010

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature:

Vincent McKie

Date: 08/05/2009

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Megaquant D-Fructose/D-Glucose Assay Kit
Cat. No: K-FRGLMQ
Lot. No: 81002
Expiry Date: December, 2010

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 10/10/2008

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Megaquant D-Fructose/D-Glucose Assay Kit
Cat. No: K-FRGLMQ
Lot. No: 61007
Expiry Date: October, 2009

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: *Patricia Rossiter*

Date: 19/10/2006

Patricia Rossiter
Quality Control Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Megaquant D-Fructose/D-Glucose Assay Kit
Cat. No: K-FRGLMQ
Lot. No: 60336
Expiry Date: March, 2009

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: *Patricia Rossiter*

Date: 30/03/2006

Patricia Rossiter
Quality Control Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)