



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Fructan Assay Kit
Cat. No: K-FRUC
Lot. No: 110519
Expiry Date: December, 2014

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 20/05/2011

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Fructan Assay Kit
Cat. No: K-FRUC
Lot. No: 110518
Expiry Date: December, 2014

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 20/05/2011

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Fructan Assay Kit
Cat. No: K-FRUC
Lot. No: 110502
Expiry Date: December, 2014

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 11/05/2011

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Certificate of Analysis

Product: Fructan Assay Kit
Cat. No: K-FRUC
Lot. No: 101106
Expiry Date: December, 2014

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature:

Vincent McKie

Date: 15/11/2010

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Fructan Assay Kit
Cat. No: K-FRUC
Lot. No: 90908
Expiry Date: December, 2014

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 11/09/2009

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Fructan Assay Kit
Cat. No: K-FRUC
Lot. No: 90123
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 23/02/2009

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Fructan Assay Kit
Cat. No: K-FRUC
Lot. No: 90122
Expiry Date: December, 2010

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 23/02/2009

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Fructan Assay Kit
Cat. No: K-FRUC
Lot. No: 80115
Expiry Date: December, 2010

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 28/01/2008

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Fructan Assay Kit
Cat. No: K-FRUC
Lot. No: 71008
Expiry Date: December, 2010

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: Vincent M^cKie

Date: 26/06/2007

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Fructan Assay Kit
Cat. No: K-FRUC
Lot. No: 61008
Expiry Date: October, 2009

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: *Patricia Rossiter*

Date: 23/10/2006

Patricia Rossiter
Quality Control Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Fructan Assay Kit
Cat. No: K-FRUC
Lot. No: 60902
Expiry Date: September, 2009

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: *Patricia Rossiter*

Date: 05/09/2006

Patricia Rossiter
Quality Control Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Fructan Assay Kit
Cat. No: K-FRUC
Lot. No: 60301
Expiry Date: March, 2009

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: *Patricia Rossiter*

Date: 03/03/2006

Patricia Rossiter
Quality Control Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)