



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Glucose HK Assay Kit (regular)
Cat. No: K-GLUHKR
Lot. No: 111114
Expiry Date: December, 2014

Bottle 1 101001
Bottle 2 110902
Bottle 3 101001a
Bottle 4 101001a

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 24/11/2011

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Glucose HK Assay Kit (regular)
Cat. No: K-GLUHKR
Lot. No: 110106
Expiry Date: December, 2014

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 07/01/2011

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Glucose HK Assay Kit (regular)
Cat. No: K-GLUHKR
Lot. No: 101002
Expiry Date: December, 2014

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 12/10/2010

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Glucose HK Assay Kit (regular)
Cat. No: K-GLUHKR
Lot. No: 100221
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 18/02/2010

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Glucose HK Assay Kit (regular)
Cat. No: K-GLUHKR
Lot. No: 100114
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 04/02/2010

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Glucose HK Assay Kit (regular)
Cat. No: K-GLUHKR
Lot. No: 90312
Expiry Date: December, 2010

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 26/03/2009

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Glucose HK Assay Kit (regular)
Cat. No: K-GLUHKR
Lot. No: 80503
Expiry Date: December, 2010

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature:

Vincent McKie

Date: 22/05/2008

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Glucose HK Assay Kit (regular)
Cat. No: K-GLUHKR
Lot. No: 71108
Expiry Date: December, 2010

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 30/11/2007

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Glucose HK Assay Kit (regular)
Cat. No: K-GLUHKR
Lot. No: 70425
Expiry Date: December, 2010

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: *Patricia Rossiter*

Date: 17/05/2006

Patricia Rossiter
Quality Control Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Glucose HK Assay Kit (regular)
Cat. No: K-GLUHKR
Lot. No: 60309
Expiry Date: March, 2009

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: *Patricia Rossiter*

Date: 08/03/2006

Patricia Rossiter
Quality Control Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)