



Megazyme International Ireland,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 110916
Expiry Date: December, 2014

Bottle 1 100601
Bottle 2 110402
Bottle 3 110301
Bottle 4 110302
Bottle 5 110201
Bottle 6 100501

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature:

Date: 28/09/2011

Vincent McKie
Quality Control Manager



Megazyme International Ireland,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 110915
Expiry Date: December, 2014

Bottle 1 100601
Bottle 2 110402
Bottle 3 110301
Bottle 4 100101
Bottle 5 110201
Bottle 6 100501

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature:

Date: 28/09/2011

Vincent McKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 110541
Expiry Date: December, 2014

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature:

Vincent McKie

Date: 20/05/2011

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 110540
Expiry Date: December, 2014

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 20/05/2011

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 101120
Expiry Date: December, 2014

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 15/11/2010

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 101115
Expiry Date: December, 2014

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 15/11/2010

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 100707
Expiry Date: December, 2013

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 07/07/2010

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 100706
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature:

Vincent McKie

Date: 07/07/2010

Vincent McKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 100705
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 07/07/2010

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 90714
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature:

Vincent McKie

Date: 08/07/2009

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 90517
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 21/05/2009

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 90516
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 21/05/2009

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 90515
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature:

Vincent McKie

Date: 21/05/2009

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 80801
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 06/08/2008

Vincent M^cKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 80112
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 18/01/2008

Vincent McKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 71005
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 18/10/2007

Vincent McKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 70911
Expiry Date: December, 2012

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 20/09/2007

Vincent McKie
Quality Control Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 70312
Expiry Date: December, 2011

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: *Patricia Rossiter*

Date: 27/03/2007

Patricia Rossiter
Quality Control Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 60807
Expiry Date: August, 2009

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: *Patricia Rossiter*

Date: 22/08/2006

Patricia Rossiter
Quality Control Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 60806
Expiry Date: August, 2009

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 22/08/2006

Patricia Rossiter
Quality Control Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 60706
Expiry Date: July, 2009

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: 

Date: 04/07/2006

Patricia Rossiter
Quality Control Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)



Megazyme International Ireland Limited,
Bray Business Park,
Bray, Co. Wicklow,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Yeast Beta-Glucan Assay Kit
Cat. No: K-YBGL
Lot. No: 60406
Expiry Date: April, 2009

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature: *Patricia Rossiter*

Date: 07/04/2006

Patricia Rossiter
Quality Control Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)