



Megazyme,

Bray Business Park,
Bray, Co. Wicklow,
A98 YV29,
Ireland.

Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Ammonia (Rapid) Assay Kit
Cat. No: K-AMIAR
Lot. No: 180811
Expiry Date: December, 2019

Bottle 1 171001
Bottle 2 160602
Bottle 3 150703g
Bottle 4 130604b

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature:  **Date:** 24/08/2018

Maciej Peplinski
Stock Co-Ordinator

p.p. Ida Lazewska QC/QA Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)