



**Megazyme International Ireland,**  
Bray Business Park,  
Bray, Co. Wicklow,  
A98 YV29,  
Ireland.  
Tel: + 353 1 2861220 Fax: + 353 1 2861264

## Certificate of Analysis

**Product:** Glucose HK Assay Kit (regular)  
**Cat. No:** K-GLUHK-220A  
**Lot. No:** 160612  
**Expiry Date:** December, 2019

Bottle 1 150701  
Bottle 2 150602  
Bottle 3 151003  
Bottle 4 141004b

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

**Signature:**

**Date:** 22/06/2016

Páraic McGeough  
Biochemical Analyst/Research Assistant

p.p. Ida Lazewska QC/QA Laboratory Manager

*(as this certificate is a computer printout, it has not been signed or dated by hand)*