

**Megazyme,**

Bray Business Park,  
Bray, Co. Wicklow,  
A98 YV29,  
Ireland.

Tel: + 353 1 2861220 Fax: + 353 1 2861264

## Certificate of Analysis

**Product:** L-Glutamic Acid Assay Kit  
**Cat. No:** K-GLUT  
**Lot. No:** 170414  
**Expiry Date:** December, 2020

Bottle 1 160701a  
Bottle 2 160502  
Bottle 3 130401a  
Bottle 4 160904  
Bottle 5 110905b

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

**Signature:**

**Date:** 27/04/2017

Ruth Ivory  
Senior Analyst

p.p. Ida Lazewska QC/QA Laboratory Manager

*(as this certificate is a computer printout, it has not been signed or dated by hand)*