Megazyme,
Bray Business Park,
Bray, Co. Wicklow,
A98 YV29,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Malt Amylase Assay Kit
Cat. No: K-MALTA
Lot. No: 220416
Expiry Date: December, 2024

Bottle 1	200802
Bottle 2	200401
Bottle 3	210101
Bottle 4	211104
Bottle 5	210105
Bottle 6	180803A

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 5%	Corresponds
Blank reaction absorbance	+/- 10% of mean blank value	Corresponds

This lot meets our specification.

Signature: Sarah Hurley **Date:** 13-Apr-2022

Sarah Hurley
Senior Biochemical QC Analyst

p.p. Ida Lazewska QC/QA Laboratory Manager

(*as this certificate is a computer printout, it has not been signed or dated by hand*)