

**Megazyme,**

Bray Business Park,
Bray, Co. Wicklow,
A98 YV29,
Ireland.

Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: D-Mannitol/L-Arabitol Assay Kit
Cat. No: K-MANOL
Lot. No: 180608
Expiry Date: December, 2020

Bottle 1 160701b
Bottle 2 170102
Bottle 3 111201e
Bottle 4 170204b

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature:

Maciej Peplinski

Date: 13/06/2018

Maciej Peplinski
Stock Co-Ordinator

p.p. Ida Lazewska QC/QA Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)