



Megazyme,
Bray Business Park,
Bray, Co. Wicklow,
A98 YV29,
Ireland.
Tel: + 353 1 2861220 Fax: + 353 1 2861264

Certificate of Analysis

Product: Trehalose Assay Kit
Cat. No: K-TREH
Lot. No: 181217
Expiry Date: December, 2023

Bottle 1 180701a
Bottle 2 180102
Bottle 3 180403
Bottle 4 170501a
Bottle 5 161106a
Bottle 6 170706a

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

Signature:  **Date:** 13/12/2018

Maciej Peplinski
Stock Co-Ordinator

p.p. Ida Lazewska QC/QA Laboratory Manager

(as this certificate is a computer printout, it has not been signed or dated by hand)