



**Megazyme International Ireland Limited,**  
Bray Business Park,  
Bray, Co. Wicklow,  
Ireland.  
Tel: + 353 1 2861220 Fax: + 353 1 2861264

## Certificate of Analysis

**Product:** Total Starch Control Flours Assay Kit  
**Cat. No:** K-TSCK  
**Lot. No:** 150404  
**Expiry Date:** December, 2020

Bottle 1: 20901a  
Bottle 2: 60901a  
Bottle 3: 21101b  
Bottle 4: 21102b  
Bottle 5: 21105b

Test	Specification	Result
Assay performance	Reaction completed within time stated in kit booklet	Corresponds
	Target value for recommended standard material +/- 2 %	Corresponds
Blank reaction absorbance	+/- 10 % of mean blank value	Corresponds

This lot meets our specification.

**Signature:** 

**Date:** 09/04/2015

Páraic McGeough  
Biochemical Analyst/Research Assistant

p.p. Ida Lazewska QC/QA Laboratory Manager

*(as this certificate is a computer printout, it has not been signed or dated by hand)*